



DEPARTMENT OF ENTOMOLOGY Phone 01425-254022 (O)
SKN COLLEGE OF AGRICULTURE
(Sri Karan Narendra Agriculture University, JOBNER)
JOBNER-3033

No. SKN/Ento./ELU/2021/29

Date: 23.01.2021

LIMITED TENDER NOTICE


Sealed tender are invited from the reputed firm to supply the following items/equipments under EL-unit-II. The tender will be accepted upto 01-02-2021 (11.00 AM) and will be opened on the same day at 12.15 PM by the committee in the Department of Agronomy. The Dean has all rights to accept/ reject any tender without assigning the reason thereof. F.O.R. at Department of Entomology, SKNCOA, Jobner. Earnest money will be deposited in the Dean Office.

| S.No. | Name of items | Quantity | Specifications |
|-------|-----------------------|----------|---|
| 1. | Refrigerator | 1 | New Modal, Type: Double door , 245 liter capacity, Frost free, 2 Star rating 2020, Digital inverter Compressor. Weight, one year warranty with 10 years on compressor. |
| 2 | High speed centrifuge | 1 | Specifications: Rotor Capacity: 12 place closed rotor with metal safety lid Maximum Speed RPM / RCF (g) : 15000 rpm / 15596 g Display : Large Backlit LCD display Imbalance detection : Imbalance Detection with auto cut – off Noise Level : < 60 dB Motor Type : (BLDC) Brushless DC motor Timer setting : 1 to 999 mins Speed / Time setting Rotors and Adaptors:- (included in standard pack) 12 place x 1.5/2.0 ml microtubes Reduction Adaptors for 0.2 ml microtubes Reduction Adaptors for 0.4/0.5 ml microtubes PCR Strip Rotor (2 x 8 x 0.2 ml with metal safety lid). |
| 3 | Homogenizer | 1 | Max AMP-0.45, Supply- 220v, 50Hz,1*AC,RPM-2800, Stirring Capacity-100ml, Shaft Diameter -8mm |

Copy to:

1. Dean, SKNCOA, Jobner
2. Convener / Member, Tender Committee
3. M/s. _____
4. The Cashier/ Accounts Section, SKNCOA, Jobner
5. CIMCA, SKNAU Jobner to upload the same on University website and <http://sppp.rajasthan.gov.in>
6. Notice board Office/ Municipal board/ Market, Jobner


HEAD
Department of Entomology
SKN College of Agriculture
JOBNER (Jobner) 303 329


HEAD
Department of Entomology
SKN College of Agriculture
JOBNER (Jobner) 303 329

सीमित बोली प्रारूप एवं अन्य शर्तें

1. सीमित बोली सूचना में प्रकाशित शर्तें राजस्थान लोक उपापन में पारदर्शिता अधिनियम, 2012, राजस्थान लोक उपापन में पारदर्शिता नियम, 2013, सामान्य वित्तीय एवं लेखा नियम में उल्लेखित नियम-68 एवं वित्त विभाग, राजस्थान की अधिसूचना दिनांक 19.11.2015 की शर्तें निविदा का भाग मानी जाएगी। आपूर्ति एवं स्थापना (Instalation) श्री कर्ण नरेन्द्र कृषि विश्वविद्यालय, जोबनेर पर करनी होगी।
2. सीमित बोली प्रपत्र के साथ संलग्न प्रपत्र 'अ' में रेफरीजरेटर, हाई स्पीड सेन्टीप्यूज एव हामोजेनाइजर हेतु बोलीदाता अपनी दरें दर्शाएँ। विश्वविद्यालय के प्रपत्र के प्रारूप के अतिरिक्त किसी अन्य दस्तावेज पर दी गई दरें मान्य नहीं होगी।
3. सीमित बोली मुहरबंद लिफाफे में प्रस्तुत की जाएगी जिस पर रेक एवं कूलर के क्रय के लिए सीमित बोली अंकित होना चाहिए।
4. फर्म द्वारा प्रस्तुत बिलों का भुगतान विश्वविद्यालय द्वारा फर्म को आदेशित रेफरीजरेटर, हाई स्पीड सेन्टीप्यूज एव हामोजेनाइजर के संतोषप्रद होने की पुष्टि संबंधित कार्यालय प्रभारी द्वारा प्रस्तुत करने पर किया जाएगा।
5. सीमित बोलीदाता को किसी भी प्रकार का कोई अग्रिम भुगतान नहीं किया जाएगा। प्रत्येक कार्यादेश/क्रयादेश का कार्य/आपूर्ति संतोषप्रद रूप से पूर्ण होने एवं विश्वविद्यालय में स्वीकार करने के बाद ही भुगतान की कार्यवाही की जाएगी।
6. सीमित बोलीदाता सीमित बोली प्रपत्र की शर्तों के प्रत्येक पृष्ठ पर मुहर सहित हस्ताक्षर करेगा तथा सीमित बोली प्रस्ताव के अंतिम पृष्ठ पर सभी शर्तों को सम्पूर्ण रूप में स्वीकार करने की सहमति देते हुए अलग से हस्ताक्षर करेगा।
7. सीमित बोली प्रपत्र में किसी प्रकार की काँट छॉट व ओवर राईटिंग नहीं होनी चाहिए। काँट छॉट/ओवर राईटिंग होने पर बोली रद्द की जा सकेगी।
8. निर्धारित समय में कार्य सम्पन्न नहीं करने/आपूर्ति नहीं करने पर राज्य सरकार के सामान्य वित्तीय एवं लेखा नियमों व राजस्थान लोक उपापन में पारदर्शिता अधिनियम, 2012 एवं नियम 2013 के अनुसार राशि वसूल की जाएगी।
9. क्रयादेश/कार्यादेश के अनुसार रेफरीजरेटर, हाई स्पीड सेन्टीप्यूज एव हामोजेनाइजर की आपूर्ति का कार्य मय स्थापना (Instalation) 03 दिवस में करना होगा।
10. सफल सीमित बोलीदाता के पूर्ण रूप से रेफरीजरेटर, हाई स्पीड सेन्टीप्यूज एव हामोजेनाइजर की आपूर्ति करने में असफल रहने पर अथवा कार्य संतोषजनक नहीं करने पर विभागाध्यक्ष, कीट विज्ञान विभाग, श्री कर्ण नरेन्द्र कृषि विश्वविद्यालय, जोबनेर, जयपुर को यह अधिकार होगा कि अनुमोदित फर्म के हर्ज खर्च पर पेनल्टी सहित अन्य व्यवस्था द्वारा आदेशित कार्य क्रय अन्य फर्म से करा कर सकेंगे।
11. समस्त सीमित बोली प्रपत्र दिनांक 01.01.2021 को 11:00 बजे तक कीट विज्ञान विभाग, श्री कर्ण नरेन्द्र कृषि महाविद्यालय, जोबनेर, जयपुर के कार्यालय में प्राप्त किए जाकर स्वीकार किए जाएंगे एवं उसी दिन 12.15 बजे शस्य विज्ञान विभाग में निविदा कमेटी के समक्ष




खोले जाएँगे। निर्धारित दिनांक व समय के पश्चात प्राप्त होने वाले सीमित बोली प्रपत्रों पर विचार नहीं किया जाएगा।

12. उपर्युक्त शर्तों के अतिरिक्त अन्य शर्तें निविदा सूचना एवं सामान्य वित्तीय एवं लेखा नियम व राजस्थान लोक उपापन में पारदर्शिता अधिनियम, 2012 एवं नियम 2013 एवं वित्त विभाग, राजस्थान की अधिसूचना दिनांक 19.11.2015 के प्रावधानानुसार लागू होगी।
13. किसी भी उत्पन्न विवाद की स्थिति में न्यायालय क्षेत्र, जयपुर (राजस्थान) होगा।



HEAD
Department of Entomology
SKN College of Agriculture
Jodhpur (Raj. Pr) 303 329



(Dr.S.K. Khinchi)
Incharge EL Unit - II

मैं/हमने उपर्युक्त सभी शर्तों को सावधानी पूर्वक पढ़ लिया है एवं समझ लिया है तथा मैं/हम उपर्युक्त समस्त शर्तों से प्रतिबंधित रहूँगा/रहेगे।

निविदादाता के हस्ताक्षर मय
मोहर

Tender format

| S. N. | Specifications | Qty. | Brand/ Make | Rate per unit including all taxes (Rs.) | Total amount (Rs.) |
|-------|--|------|-------------|---|--------------------|
| 1. | Refrigerator New Modal. Type: Double door , 245 liter capacity, Frost free, 2 Star rating 2020. Digital inverter Compressor. Weight, one year warranty with 10 years on compressor. | 1 | | | |
| 2 | High speed centrifuge: Rotor Capacity: 12 place closed rotor with metal safety lid Maximum Speed RPM / RCF (g) : 15000 rpm / 15596 g Display : Large Backlit LCD display Imbalance detection : Imbalance Detection with auto cut – off Noise Level : < 60 dB Motor Type : (BLDC) Brushless DC motor Timer setting : 1 to 999 mins Speed / Time setting Rotors and Adaptors:- (included in standard pack) 12 place x 1.5/2.0 ml microtubes Reduction Adaptors for 0.2 ml microtubes Reduction Adaptors for 0.4/0.5 ml microtubes PCR Strip Rotor (2 x 8 x 0.2 ml with metal safety lid). | 1 | | | |
| 3 | Homogenizer Max AMP-0.45, Supply- 220v, 50Hz, 1*AC, RPM-2800, Stirring Capacity-100ml. Shaft Diameter -8mm | 1 | | | |

निविदादाता के हस्ताक्षर मय
मोहर